

---

# **Bridging for Drug-Device and Biologic-Device Combination Products Guidance for Industry**

## ***DRAFT GUIDANCE***

**This guidance document is being distributed for comment purposes only.**

Comments and suggestions regarding this draft document should be submitted within 60 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. All comments should be identified with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions regarding this draft document, contact (CDER) Irene Chan at 301-796-3962 or Robert Berlin at 301-796-8828, (CBER) Office of Communication, Outreach, and Development at 240-402-8010, (CDRH) CDRH product jurisdiction officer at [CDRHProductJurisdiction@fda.hhs.gov](mailto:CDRHProductJurisdiction@fda.hhs.gov), or (OCP) Patricia Love at [patricia.love@fda.hhs.gov](mailto:patricia.love@fda.hhs.gov).

**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Drug Evaluation and Research (CDER)  
Center for Biologics Evaluation and Research (CBER)  
Center for Devices and Radiological Health (CDRH)**

**December 2019  
Combination Products**

---

# Bridging for Drug-Device and Biologic-Device Combination Products Guidance for Industry

*Additional copies are available from:*

*Office of Communications, Division of Drug Information  
Center for Drug Evaluation and Research  
Food and Drug Administration  
10001 New Hampshire Ave., Hillandale Bldg., 4th Floor  
Silver Spring, MD 20993-0002*

*Phone: 855-543-3784 or 301-796-3400; Fax: 301-431-6353; Email: [druginfo@fda.hhs.gov](mailto:druginfo@fda.hhs.gov)  
<https://www.fda.gov/drugs/guidance-compliance-regulatory-information/guidances-drugs>*

*or*

*Office of Communication, Outreach, and Development  
Center for Biologics Evaluation and Research  
Food and Drug Administration  
10903 New Hampshire Ave., Bldg. 71, Room 3128  
Silver Spring, MD 20993-0002*

*Phone: 800-835-4709 or 240-402-8010; Email: [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov)  
<https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>*

*or*

*Office of Policy  
Center for Devices and Radiological Health  
Food and Drug Administration  
10903 New Hampshire Ave., Bldg. 66, Room 5431  
Silver Spring, MD 20993-0002*

*Email: [CDRH-Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov)  
<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/guidance-documents-medical-devices-and-radiation-emitting-products>*

**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Drug Evaluation and Research (CDER)  
Center for Biologics Evaluation and Research (CBER)  
Center for Devices and Radiological Health (CDRH)**

**December 2019  
Combination Products**

---

*Contains Nonbinding Recommendations*

*Draft—Not for Implementation*

**TABLE OF CONTENTS**

<b>I.</b>	<b>INTRODUCTION.....</b>	<b>1</b>
<b>II.</b>	<b>BACKGROUND .....</b>	<b>3</b>
<b>III.</b>	<b>DEVELOPING AN ANALYTICAL FRAMEWORK FOR IDENTIFYING INFORMATION GAPS TO INFORM A BRIDGING AND LEVERAGING APPROACH.....</b>	<b>4</b>
<b>IV.</b>	<b>BRIDGING AND LEVERAGING EXAMPLES.....</b>	<b>6</b>
	<b>A. Bridging Within an IND from a Drug Developed in a Prefilled Syringe to a Drug Developed in an Autoinjector .....</b>	<b>7</b>
	<b>B. Bridging From One Autoinjector (Prototype 1) to Another Autoinjector (Prototype 2) for the Same Drug; After Phase 3 Studies Have Been Completed but Before NDA Submission .....</b>	<b>11</b>
	<b>C. Bridging of Data From Combination Product That Employs the Same Device Combined With a Different Drug .....</b>	<b>13</b>

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27

# **Bridging for Drug-Device and Biologic-Device Combination Products Guidance for Industry<sup>1</sup>**

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff responsible for this guidance as listed on the title page.

## **I. INTRODUCTION**

This guidance provides recommendations to industry and FDA staff on how to approach bridging in new drug applications (NDAs) or biologics license applications (BLAs) for drug-device and biologic-device single entity or copackaged combination products including the following:<sup>2</sup>

- Bridging of information related to a combination product that employs a different device constituent part or parts<sup>3</sup> with the same drug constituent part or parts<sup>4</sup> as the proposed combination product

---

<sup>1</sup> This guidance has been prepared by the Office of Surveillance and Epidemiology and the Office of New Drugs in the Center for Drug Evaluation and Research, in cooperation with the Center for Biologics Evaluation and Research, the Center for Devices and Radiological Health, and the Office of Combination Products at the Food and Drug Administration. This guidance is one of several documents FDA is issuing to fulfill the performance goals under the sixth authorization of the Prescription Drug User Fee Act (PDUFA VI).

<sup>2</sup> See section 503(g) of the Federal Food, Drug, and Cosmetic Act (FD&C Act) and the definition of combination products in 21 CFR 3.2.

<sup>3</sup> See constituent part definition in 21 CFR 4.2.

<sup>4</sup> For purposes of this guidance, except where specifically indicated, references to *drug* or *drug constituent part or parts* include a drug or biological product constituent part submitted as part of a combination product for approval or approved under section 505(c) of the FD&C Act (21 U.S.C. 355(c)) or licensed under section 351(a) of the Public Health Service Act (PHS Act). Some of the principles applicable to products submitted for approval under section 505(b) of the FD&C Act or licensure under section 351(a) of the PHS Act may also be applicable to products submitted for approval under section 505(j) of the FD&C Act or licensure under section 351(k) of the PHS Act. In addition, the scientific principles discussed in this guidance may be applicable to combination product submissions under sections 515, 513(f)(2), or 510(k) of the FD&C Act.

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

- 28       • Bridging of information related to a combination product that employs a different drug  
29       constituent part or parts with the same device constituent part or parts as the proposed  
30       combination product

31  
32 For the purposes of this guidance, the term *bridging* refers to the process of establishing the  
33 scientific relevance of information developed in an earlier phase of the development program or  
34 another development program to support the combination product for which an applicant is  
35 seeking approval. Once the applicant has established the relevance of such information to (i.e.,  
36 bridged to) its product, the applicant can leverage that information to streamline its development  
37 program.<sup>5</sup> From a scientific perspective, an applicant must bridge its current application to  
38 information developed in an earlier phase of the development program or another development  
39 program if the applicant wishes to leverage that information in its current application. For  
40 certain types of applications, the use of information from another development program may  
41 require that the applicant own the information or have a right of reference.<sup>6</sup>

42  
43 With respect to the recommendations and examples in this guidance, it is assumed that the  
44 applicant owns or has a right of reference or use that allows the applicant to use information  
45 from another development program.

46  
47 This guidance seeks to clarify how to bridge to information gathered from another development  
48 program to leverage that information in support of an application. To facilitate that process, this  
49 guidance describes an approach for an applicant to identify and address information gaps for an  
50 application.

51  
52 This guidance applies to the following:

- 53  
54       • Human prescription combination products that are the subject of an investigational new  
55       drug application (IND) under 21 CFR part 312, an NDA under 21 CFR part 314, or a  
56       BLA under 21 CFR part 601
- 57  
58       • Human nonprescription combination products that are the subject of an IND, NDA, or  
59       BLA (as opposed to those covered in a final or tentative over-the-counter drug  
60       monograph)

61

---

<sup>5</sup> There are certain regulatory considerations that apply to reliance on certain types of information in certain applications (e.g., reliance on a previous finding of safety and effectiveness for a drug the applicant does not own or to which it has no right of reference in a 505(b)(2) application) but discussion of those considerations is beyond the scope of this guidance. See the draft guidance for industry *Applications Covered by Section 505(b)(2)* (October 1999). When final, this guidance will represent the FDA’s current thinking on this topic. For the most recent version of a guidance, check the FDA guidance web page at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents>.

<sup>6</sup> See, for example, 21 CFR 314.3, “Right of reference or use is the authority to rely upon, and otherwise use, an investigation for the purpose of obtaining approval of an NDA [new drug application], including the ability to make available the underlying raw data from the investigation for FDA audit, if necessary.”

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

62 All such products in this guidance are referred to as combination products. Except where it is  
63 specifically indicated that this is not the case, the terms *drug* and *drug constituent part* are used  
64 interchangeably and also refer to biological products and biological product constituent parts; the  
65 terms *device* and *device constituent part* are used interchangeably, and persons responsible for  
66 product development are referred to as *applicants*.

67  
68 Although this guidance is intended to help applicants consider the type and scope of information  
69 that may be leveraged for a combination product development program, this guidance does not  
70 address all of the issues applicable to any particular combination product. The Agency  
71 encourages applicants to contact FDA to discuss specific information needed to support their  
72 individual applications.<sup>7</sup>

73  
74 This guidance does not discuss the appropriate regulatory pathway an applicant should use to  
75 bring a particular combination product to market.<sup>8</sup>

76  
77 In general, FDA’s guidance documents do not establish legally enforceable responsibilities.  
78 Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only  
79 as recommendations, unless specific regulatory or statutory requirements are cited. The use of  
80 the word *should* in Agency guidances means that something is suggested or recommended, but  
81 not required.

82  
83  
84 **II. BACKGROUND**

85  
86 While drugs, devices, and biological products retain their discrete regulatory identities when they  
87 are constituent parts of a combination product, combination products comprise a distinct  
88 category of medical products that can be subject to specialized regulatory requirements.<sup>9</sup>  
89 Accordingly, the regulatory requirements for combination products arise from the statutory and  
90 regulatory requirements applicable to drugs, devices, and biological products.<sup>10</sup> Consistent with  
91 section 503(g) of the Federal Food, Drug, and Cosmetic Act, FDA is committed to applying a  
92 consistent, risk-based approach to address similar regulatory questions, including scientific  
93 questions, using relevant expertise from the lead and consulted centers within FDA.

---

<sup>7</sup> See the draft guidance for industry *Formal Meetings Between the FDA and Sponsors or Applicants of PDUFA Products* (December 2017). When final, this guidance will represent FDA’s current thinking on this topic.

<sup>8</sup> See the draft guidance for industry and FDA staff *Principles of Premarket Pathways for Combination Products* (February 2019). When final, this guidance will represent FDA’s current thinking on this topic.

<sup>9</sup> See the final rule, “Current Good Manufacturing Practice Requirements for Combination Products,” published January 22, 2013 (21 CFR Part 4, Subpart A; 78 FR 4307-22) and the final rule, “Postmarketing Safety Reporting for Combination Products,” published December 20, 2016 (21 CFR Part 4, Subpart B; 81 FR 92603-26).

<sup>10</sup> *Ibid.*

## *Contains Nonbinding Recommendations*

### *Draft—Not for Implementation*

95 Depending on an applicant’s development program, there may be circumstances in which an  
96 applicant has its own existing information (or rights of reference to information) about another  
97 combination product or a proposed constituent part that may be leveraged to support approval of  
98 the proposed combination product if an appropriate bridge can be established. In general, FDA  
99 would require additional data and information only if the information were needed to address  
100 additional questions of safety or effectiveness raised by the proposed use or function of a  
101 constituent part in the combination product.<sup>11</sup> For example, in general, if a stand-alone device  
102 proposed to be used as a device constituent part of a combination product has been previously  
103 approved or cleared, the applicant may be able to leverage relevant existing device-related data,  
104 provided that the data has been bridged (i.e., shown to be scientifically relevant), for the  
105 development of a new combination product.

106  
107 In some cases, the amount of information that can be leveraged for such proposed combination  
108 products may be minimal, or leveraging may not be possible. For example, a change in route of  
109 administration for a complex biological product may raise additional safety and/or efficacy  
110 considerations, and such considerations may make it difficult to bridge to the proposed  
111 combination product. Discussions with the Agency about planned leveraging are appropriate to  
112 identify questions early in drug development.

113  
114

### 115 **III. DEVELOPING AN ANALYTICAL FRAMEWORK FOR IDENTIFYING** 116 **INFORMATION GAPS TO INFORM A BRIDGING AND LEVERAGING** 117 **APPROACH**

118  
119 Developing a framework that identifies where information gaps may exist in a combination  
120 product development program is an important task for applicants. The following information  
121 assumes that applicants are familiar with existing FDA regulations, guidance documents, and  
122 resources on drug and device development available from the Center for Drug Evaluation and  
123 Research (CDER), the Center for Biologics Evaluation and Research, the Center for Devices and  
124 Radiological Health, and the Office of Combination Products to assess the information that  
125 should be included in an IND, NDA, or BLA, as appropriate. Under this premise, the example of  
126 a framework below supposes that an applicant seeks to bridge from the FDA-approved, drug-  
127 device (delivery system) Combination Product A to proposed Combination Product B. The  
128 Agency recommends that an applicant use the stepwise approach presented below to conduct a  
129 gap analysis for the proposed Combination Product B:

130

131 **Step 1.** Identify all differences between Combination Products A and B, and consider the  
132 potential effect of the individual and aggregate differences on the safety and effectiveness  
133 profile for Combination Product B.

134

---

<sup>11</sup> See section 503(g)(3) of the FD&C Act: “The [FDA] may require that the sponsor of such combination product submit to the [FDA] only data or information that the [FDA] determines is necessary to meet the standard for clearance or approval, as applicable, under this Act or the Public Health Service Act, including any incremental risks and benefits posed by such combination product, using a risk-based approach and taking into account any prior finding of safety and effectiveness or substantial equivalence for the approved constituent part relied upon by the applicant.”

## ***Contains Nonbinding Recommendations***

### ***Draft—Not for Implementation***

135 Specifically, for Combination Product B in comparison with Combination Product A, the  
136 safety and effectiveness profile should include a clear, comprehensive listing of the  
137 differences in the device constituent part, the drug constituent part, and the combination  
138 product as a whole. Some examples of the potential effect of a change in drug or device  
139 constituent part for Combination Product B compared to the existing safety and/or  
140 effectiveness profile for Combination Product A include the following:

- 141
- 142 • Changes to the local injection adverse reaction profile including those related to an  
143 increase in drug concentration, a change in drug viscosity or formulation, or a change  
144 in injection rate
- 145
- 146 • Change in the dose accuracy of the same device constituent part when the drug  
147 formulation is changed
- 148
- 149 • Changes in the manufacturing process and/or device constituent part that may affect  
150 drug quality
- 151
- 152 • Change in whether the intended users can use Combination Product B safely and  
153 effectively when the user interface of the device constituent part changes
- 154
- 155 • Changes in the bioavailability of the drug and/or its metabolic profile that can occur  
156 because of changes in the device, formulation, or route of administration, such as the  
157 following:
  - 158 – Changes in the needle depth, tissue plane, or rate of infusion
  - 159
  - 160
  - 161 – Change in drug formulation that results in differential lung depositions even when  
162 the drug is administered with the same device
  - 163
- 164 • Changes in drug formulation that can affect the leachable and extractable profiles of  
165 the combination product
- 166

167 **Step 2.** Identify existing information for Combination Product B (i.e., information that has  
168 been gathered or generated through studies and assessments of the proposed combination  
169 product itself) and compare it to the safety and effectiveness submission requirements  
170 necessary for approval.

171

172 **Step 3.** Identify and explain how and why existing information on Combination Product A  
173 can be bridged and leveraged to support approval of Combination Product B, taking into  
174 account the considerations in Step 1 and the information already gathered in Step 2.

175

176 **Step 4.** Focus on any information gaps remaining from Steps 2 and 3, and consider whether  
177 other existing information, outside of that directly gathered for Combination Product A or B,

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

178 can be reviewed and used to address these gaps under the proposed regulatory pathway as  
179 described in Step 5 below.<sup>12</sup>

180  
181 **Step 5.** Compare findings from Step 2 through 4 and identify the remaining gaps in  
182 information that need to be addressed in the product application.

183  
184 After completing a gap analysis, FDA recommends that applicants meet with FDA’s lead center  
185 review division along with consulting reviewers to discuss what new information or studies may  
186 be needed to support the application for Combination Product B.

187  
188 **Special considerations:** The stepwise framework and associated analyses described above  
189 represent general considerations regarding how the applicant should prepare an application.  
190 However, leveraging may be challenging or not possible with some combination products  
191 because they contain complex constituent parts and/or are likely to be affected by seemingly  
192 minor changes. For example, combination products that include certain biological products or  
193 complex delivery systems may not allow the same degree of leveraging as would be possible for  
194 a combination product that includes a well-characterized drug or a well-understood device.  
195 Nonetheless, the framework and associated analyses in this guidance are at least potentially  
196 applicable to such combination products.

197  
198  
199 **IV. BRIDGING AND LEVERAGING EXAMPLES**

200  
201 In this section, we present three case examples to illustrate how an applicant might appropriately  
202 apply the above stepwise framework and associated analyses to determine the bridging strategy  
203 and informational needs in a development program, which it would then present to FDA. It is  
204 important to note that the cases represent hypothetical examples. The approach taken provides  
205 one acceptable way to break down the thought process around preparing applications from an  
206 applicant’s viewpoint. We recognize that many applicants would likely be considering multiple  
207 steps from the framework simultaneously. Most importantly, these considerations and  
208 recommendations are not intended to apply to any particular development program. Product-  
209 specific considerations will lead to differing informational requirements by FDA. We encourage  
210 applicants to discuss their particular development program and bridging strategy with FDA.

211  
212 If the applicant determines early in a drug development program that the intent will be to market  
213 multiple presentations or a presentation that is different from that studied in early development,<sup>13</sup>  
214 FDA encourages the applicant to conduct clinical studies using the device constituent parts with  
215 which it intends to market the combination product (i.e., the final finished combination product).  
216 By doing so, bridging to clinical data likely would not be needed because the data would have  
217 been developed with the final finished combination product.

---

<sup>12</sup> Note that use of certain sources of information may not be permitted under certain regulatory pathways, but that discussion is beyond the scope of this guidance.

<sup>13</sup> For the purposes of this guidance, the term *presentation* refers to the device constituent part of the combination product.

*Contains Nonbinding Recommendations*

*Draft—Not for Implementation*

219           A.       **Bridging Within an IND from a Drug Developed in a Prefilled Syringe to a**  
220                           **Drug Developed in an Autoinjector**

221  
222 In this hypothetical case example, the applicant is developing a combination product containing a  
223 new molecular entity (NME) drug constituent part with the initial plan to market it in a prefilled  
224 syringe (PFS) presentation, intended for home use by laypersons, including patients. During the  
225 course of development, the applicant decided that it would also like to market the NME in an  
226 autoinjector presentation. The final finished combination product for the newly proposed  
227 presentation will be an autoinjector assembled around the original PFS. The primary container  
228 closure in direct contact with the drug (i.e., barrel, plunger, and needle) remains the same and the  
229 drug formulation remains the same. The route of administration (subcutaneous) is the same.  
230 The applicant intends to market both the PFS and autoinjector presentations commercially.

231  
232 Using the stepwise framework, the applicant’s gap analysis identifies the following:

233  
234       **Step 1.** The applicant identifies the differences between the first and second presentations.  
235 The principal difference is the change of the device constituent part made by adding an  
236 autoinjector to the PFS combination product. In this case, the autoinjector results in three  
237 key changes: 1) it adds a new secondary container closure, 2) it changes the method of  
238 injecting the drug constituent part, and 3) it has a different user interface.

239  
240 In considering the potential effect of the individual and aggregate differences on the safety  
241 and effectiveness of the autoinjector combination product as a whole, the applicant identifies  
242 the following gap-analysis considerations:

- 243
- 244       • The difference in the user interface leaves an unanswered question regarding whether  
245 the user interface design supports safe and effective use, which may change its safety  
246 and effectiveness profile as compared to the PFS.
  - 247
  - 248       • Changes in assembly of the prefilled syringe into the autoinjector could change  
249 quality considerations for the drug constituent part if the manufacturing process for  
250 the autoinjector adversely affects the drug, including, but not limited to, degradation  
251 associated with assembly and the effect of the process on sterility. Likewise,  
252 chemistry, manufacturing, and controls (CMC) considerations in this context include  
253 impacts on syringe resistance to breakage, functionality throughout shelf life, and  
254 expiration dating.
  - 255
  - 256       • Changes in the method of injecting the drug constituent part may affect the local  
257 adverse reaction profile because of a change in the rate at which the drug is delivered  
258 to the target tissue. For example, it is expected that the PFS is associated with more  
259 variability in injection time with real world use, whereas the autoinjector is designed  
260 to meet a specific injection time specification.
  - 261
  - 262       • Changes in the method of injecting the drug constituent part may affect the  
263 pharmacokinetic (PK) profile of the drug. For example, changing the delivery  
264 method may lead to differences including consistency or variability of injection angle,

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

265 tissue depth (potentially associated with the rate of drug delivery as determined by the  
266 injection time), and completeness of the injection.

- 267
- 268 • Assembling the autoinjector around the PFS will require, among other things,  
269 assessment of design features of the additional autoinjector combination product.<sup>14</sup>

270

271 **Step 2.** The applicant has not yet developed information specifically for the drug combined  
272 with the autoinjector presentation and, therefore, will have to either leverage existing  
273 information or develop new supporting data.

274

275 **Step 3.** The applicant conducted phase 3 studies with the PFS presentation. These studies  
276 provided data on PK, nonclinical data, toxicity, safety and effectiveness, and leachable and  
277 extractable profiles. The applicant identifies the following information that could be applied  
278 to the new autoinjector combination product and the associated rationale:

- 279
- 280 • Because the proposed drug, indication, dosage, formulation, and route of  
281 administration are the same, the applicant believes that if the PK profile is shown to  
282 be the same through testing, then nonclinical, toxicity, and safety and effectiveness

---

<sup>14</sup> Combination products are subject to 21 CFR part 4, which sets forth current good manufacturing practice (CGMP) requirements for combination products. The constituent parts of a combination product retain their regulatory status (as a drug or device, for example) after they are combined. The CGMP requirements that apply to each of the constituent parts apply to the combination product they constitute.

For single-entity and co-packaged combination products that include both a drug and a device, such as those covered in this guidance, manufacturers may implement a streamlined approach for these combination products (21 CFR 4.4(b)). Under this approach, combination product manufacturers may meet the requirements of both the drug CGMPs and device quality system QS regulation by designing and implementing a CGMP operating system that demonstrates compliance with the drug CGMPs and the following provisions from the device QS regulation: 21 CFR 820.20 (management responsibility); 21 CFR 820.30 (design controls); 21 CFR 820.50 (purchasing controls); 21 CFR 820.100 (corrective and preventive action); 21 CFR 820.170 (installation); and 21 CFR 820.200 (servicing). See 21 CFR 4.4(b)(1). One of the specified QS regulation provisions codifies the obligation to comply with 21 CFR 820.30 design controls requirements for these drug-device combination products, including design verification and validation. See 21 CFR 4.4(b)(1)(ii). Design control activities confirm that there are no negative interactions between constituent parts and ensure that their combined use results in a combination product that is safe and effective and performs as expected. The focus of design control discussion in this guidance is the information required to demonstrate that the final combination product achieves its identified performance targets under the identified conditions of use, as opposed to the procedural requirements of 21 CFR 820.30 for developing and managing such information (e.g., requirements concerning design and development planning and design history file). Data needed to make such design verification and validation demonstrations vary depending on the combination product and its intended use but typically include, among other things, bench data, preclinical/clinical testing data, and human factors (HF) studies. For further information on design control requirements for combination products, see the guidance for industry and FDA staff *Current Good Manufacturing Practice Requirements for Combination Products* (January 2017) (2017 CGMP Guidance for Combination Products).

Moreover, a biological product regulated under section 351 of the Public Health Service Act is also, by definition, a drug or a device. Accordingly, for combination products that include a biological product, in addition to complying with the drug CGMP and device QS regulation requirements as applicable in accordance with 21 CFR part 4, manufacturers of such products must comply with the CGMP requirements in 21 CFR parts 600 through 680 that would apply to the biological product if it were not part of a combination product (21 CFR 4.4(b)(3)).

## *Contains Nonbinding Recommendations*

### *Draft—Not for Implementation*

283 data gathered in the existing clinical program for the PFS presentation could also  
284 apply for the autoinjector presentation.

285

- 286 • Because the primary container closure (the PFS) is the same for both presentations  
287 (drug will be in contact with the glass of the PFS, elastomeric plunger, and needle  
288 with use), and because the applicant expects that the secondary container closure  
289 autoinjector materials, during manufacture and storage of the combination product,  
290 will not come in direct contact with the drug or change its characteristics, the  
291 applicant believes that the leachable and extractable profile gathered for the PFS  
292 presentation should also apply for the autoinjector presentation.

293

294 **Step 4.** The applicant considers whether other existing information may be leveraged to  
295 support the items in Step 4. In the course of examining other available information, the  
296 applicant identifies the following:

297

298 An autoinjector with the same user interface was previously approved as part of a  
299 combination product with another drug in the applicant's portfolio. The approved  
300 combination product was developed for a different disease state and indication and for use in  
301 a different patient population with differing injection sites. The approved combination  
302 product has been marketed for two years and there are currently no adverse compliance  
303 actions or postmarketing safety issues under investigation by the applicant. The applicant  
304 considers whether it is appropriate to establish a bridge between the previously approved  
305 combination product and the proposed autoinjector in order to leverage human factors  
306 validation data. The applicant recognizes, however, that since the product was developed for  
307 another population and indication, it will be challenging to bridge the applications and  
308 intends to conduct a HF validation study and prepare a HF validation study report to be  
309 submitted as part of the marketing application.<sup>15</sup>

310

311 The applicant also believes that it should be able to bridge the autoinjectors with regard to  
312 device performance that is unrelated to the drug since this would be unchanged between the  
313 products. In particular, because the previously approved autoinjector design used the same  
314 syringe with the same prestaked needle for its prefilled drug, the applicant believes it can  
315 leverage design verification data unrelated to the drug being injected (e.g., extended needle  
316 length, autoinjector activation force, and cap removal force). The applicant, however,  
317 intends to generate additional verification data on factors affected by the drug (e.g., dose  
318 accuracy, injection time, etc.). Additionally, the applicant considers the possibility that the  
319 change in indication, injection site or user population could impact the acceptability, from a  
320 validation perspective, of dose accuracy, extended needle length, injection time, autoinjector  
321 activation force, cap removal force and other autoinjector performance specifications.  
322 Therefore, the applicant plans to provide design validation confirming that the autoinjector  
323 performance specifications are adequate for the new drug.

---

<sup>15</sup> Regardless of whether HF studies are submitted for the marketing application, such studies and/or analysis of whether the studies are needed may still need to be included as part of design control documentation for the combination product. See, for example, 21 CFR 820.30(g) & (j).

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

324  
325  
326  
327  
328  
329  
330  
331  
332  
333  
334  
335  
336  
337  
338  
339  
340  
341  
342  
343  
344  
345  
346  
347  
348  
349  
350  
351  
352  
353  
354  
355  
356  
357

**Step 5.** The applicant determines that the following information may still be needed:

- Human factors (HF) validation data for the autoinjector presentation to support the new user interface.
- Local adverse event data to assess how the new interface may affect pain on delivery. This may include any potential change to injection time, which may change the rate at which the drug is delivered to the target tissue.
- CMC and engineering data: syringe resistance to breakage, functionality, maintenance of sterility over shelf life, degradation of the drug from the assembly process, and expiration dating. Note that the company’s testing related to sterility and degradation would primarily be intended to verify that the new process did not create issues.
- Design verification and validation data for the autoinjector presentation,<sup>16</sup> including dose accuracy and injection time.<sup>17</sup> The applicant also intends to provide a copy of design control documentation for the delivery system and combination product as a whole, including documentation of design requirements and specifications, design verification, design validation, and risk analysis for use of the applicant’s previously approved autoinjector with a new drug.
- The applicant intends to assess how any changes in drug delivery affect the PK profile of the combination product. Changes in delivery include changes in the tissue plane in which the drug is delivered, changes to the rate of delivery (because of change in injection time between the PFS and autoinjector), and changes in the consistency of the injection angle. The extent to which existing safety data or effectiveness data can be bridged and leveraged will depend on PK comparisons that will allow assessment of any differences in bioavailability between the products. If no differences are observed in the PK profile the applicant will leverage nonclinical, toxicity, and safety and effectiveness data gathered in the existing clinical program for the PFS presentation. If differences are observed in the PK profile between the two presentations (e.g., in maximum concentration, in area under the curve, in shape

---

<sup>16</sup> See, for example, 21 CFR 820.30(f)-(g). Design verification confirms that the combination product meets the applicant’s design requirements/specifications (21 CFR 820.30(f)); see also 2017 CGMP Guidance for Combination Products, at 23. Design verification activities may include, for example, performance tests, safety tests, or visual inspections (2017 CGMP Guidance for Combination Products, at 23). Design validation ensures that the combination product is designed correctly to achieve its intended purpose(s) (21 CFR 820.30(g)). Design validation may include simulated use testing or clinical/nonclinical evaluation, including HF and software validation (2017 CGMP Guidance for Combination Products, at 24).

<sup>17</sup> See the guidance for industry and FDA staff *Technical Considerations for Pen, Jet, and Related Injectors Intended for Use with Drugs and Biological Products* (June 2013).

***Contains Nonbinding Recommendations***

*Draft—Not for Implementation*

358 of the concentration-time profile), the applicant intends to gather additional  
359 information to evaluate clinical effect of these differences.<sup>18</sup>

360  
361

**B. Bridging From One Autoinjector (Prototype 1) to Another Autoinjector  
(Prototype 2) for the Same Drug; After Phase 3 Studies Have Been Completed  
but Before NDA Submission**

365

366 In this hypothetical case example, the applicant developed an autoinjector for Drug Product X,  
367 which was used through completion of its phase 3 clinical studies. During the course of  
368 development for submission, the applicant decided to make modifications to the TBM  
369 autoinjector (Prototype 2) to improve its functionality; however, the applicant does not intend to  
370 modify the device performance specifications (e.g., dose accuracy, injection time). In this case,  
371 the primary container closure in direct contact with the drug (i.e., barrel, plunger, and needle)  
372 remains the same. The assembly process used during manufacture to form the autoinjector  
373 remains the same. The drug formulation remains the same. The route of administration is the  
374 same (subcutaneous). The user interface is the same.

375

376 Using the stepwise framework, the applicant’s gap analysis identifies the following:

377

**Step 1.** The applicant identifies the differences between the clinically studied autoinjector  
(Prototype 1) presentation and the TBM autoinjector (Prototype 2) presentation. The  
dimensions and materials of the internal components of the rear and front shell subassemblies  
of the combination product were modified to improve the functionality for the combination  
product without changing the user interface.

383

In considering the potential effect of the individual and aggregate differences on safety and  
effectiveness introduced by the device modifications, the applicant identifies the following  
considerations:

387

- The revision of the autoinjector would not be expected to change the quality  
considerations for the drug constituent part if the container closure in direct contact  
with the drug and the formulation remains the same. The manufacturing process for  
Prototype 2 is comparable to that for Prototype 1, so the manufacturing process for  
the revision of the autoinjector would not be expected to affect the quality of the drug  
constituent part. However, differences in functional performance of the device  
constituent part, if any, may affect the drug constituent part.
  
- Device changes may affect the PK profile of the drug. For example, changing the  
dimensions and materials of the rear and front shell subassemblies’ internal  
components may lead to differences including consistency or variability of injection  
angle, tissue depth (potentially associated with the rate of drug delivery as  
determined by the injection time), and completeness of the injection.

400

---

<sup>18</sup> For further information, see the draft guidance for industry *Bioavailability Studies Submitted in NDAs or INDs — General Considerations* (February 2019). When final, this guidance will represent the FDA’s current thinking on this topic.

***Contains Nonbinding Recommendations***  
***Draft—Not for Implementation***

401  
402 **Step 2.** The applicant did not gather new clinical data for the TBM product, but as reflected  
403 in Step 3 conducted verification testing on the TBM autoinjector (Prototype 2) to confirm  
404 that device performance remains unchanged between the clinically studied (Prototype 1) and  
405 TBM (Prototype 2) versions of the device. This included testing in accordance with relevant  
406 standards as well as an assessment of certain performance requirements that the applicant has  
407 identified as potentially affected by the modifications to the components and that, if affected,  
408 could adversely affect the device’s operations. In this case, the following performance  
409 requirements were included in the verification testing of the TBM device (Prototype 2):  
410

- 411 • Dose accuracy
- 412 • Injection depth (needle extension)
- 413 • Injection time
- 414 • Activation force

415  
416 The applicant is aware that the above are examples of factors that could affect the drug  
417 delivery and should be assessed over combination product shelf life. The applicant has  
418 compared the identified performance requirements for both autoinjector prototypes (1 and 2)  
419 and determined that the performance requirements remain unchanged between the clinically  
420 studied (Prototype 1) and TBM (Prototype 2) versions of the device.  
421

422 **Step 3.** The applicant previously conducted phase 3 studies with Prototype 1 of the  
423 autoinjector presentation. The applicant has also performed design verification testing and  
424 completed HF validation testing on the clinically studied autoinjector (Prototype 1). The  
425 applicant identifies the following information that can be applied to the modified TBM  
426 autoinjector (Prototype 2) presentation:  
427

- 428 • As noted above, testing confirmed that the dose accuracy, delivery time, injection  
429 depth, injection angle and site of injection are the same for Prototype 1 and TBM  
430 Prototype 2 bridging that information for these presentations; therefore, the applicant  
431 has determined that the PK studies conducted using Prototype 1 of the autoinjector  
432 can be leveraged.
- 433
- 434 • The user interface is not changing. Additionally, the activation force and injection  
435 time remain the same; therefore, the applicant has determined that the HF data  
436 between Prototype 1 and 2 can be bridged. Accordingly, the HF data collected for the  
437 combination product using the Prototype 1 autoinjector can be leveraged for the  
438 combination product using the Prototype 2 autoinjector presentation.
- 439
- 440 • The proposed indication, dosage, and administration are the same and, as noted  
441 previously, there is no change to the delivery of the drug (e.g., dose accuracy,  
442 injection time, injection depth). Therefore, the applicant has determined that the  
443 nonclinical, toxicity, and safety and effectiveness data gathered in the existing clinical  
444 program using the Prototype 1 autoinjector presentation can be bridged to the  
445 Prototype 2 autoinjector presentation and leveraged.
- 446

## ***Contains Nonbinding Recommendations***

### ***Draft—Not for Implementation***

- 447
- 448
- 449
- 450
- 451
- 452
- 453
- 454
- The primary container closure is the same for both Prototype 1 and Prototype 2 autoinjectors, and the applicant can demonstrate that the autoinjector components and materials used during manufacture and storage of the combination product do not come in direct contact with the drug. Therefore, the applicant has determined that the CMC information for the drug constituent part and leachable profile gathered using the Prototype 1 autoinjector presentation can be bridged to the Prototype 2 autoinjector presentation and leveraged.
  - The primary container closure remains the same and the injection time remains unchanged from the clinically studied Prototype 1 autoinjector to the TBM Prototype 2 autoinjector. Therefore, the applicant has determined that it is possible to bridge information regarding the products and leverage the drug-device compatibility study from Prototype 1.

461 **Step 4.** The applicant determines that there is no other existing information that may be leveraged.

463

464 **Step 5.** The applicant believes no new information needs to be generated beyond that described above. The applicant intends to support the assessment through submission of data demonstrating comparability between the designs, including through submission of full design verification data for the TBM Prototype 2 to demonstrate that device performance is comparable to Prototype 1. In addition, the applicant intends to include a side-by-side comparison of the user interface for the combination product using Prototype 1, which was evaluated in the HF validation study, and the combination product using Prototype 2 as part of the NDA submission to facilitate review to demonstrate that there are no differences in the user interface.

473

474

#### **C. Bridging of Data From Combination Product That Employs the Same Device Combined With a Different Drug**

476

477

478 In this hypothetical case example, the applicant previously developed Combination Product A, which was approved by FDA in an NDA and includes a prefilled drug cartridge attached to a metered-dose inhaler. Combination Product A is indicated for the prevention and relief of bronchospasm in patients 18 years of age and older with reversible obstructive airway disease. The applicant is now early in the development of Combination Product B, which combines an NME drug constituent part with the same metered-dose inhaler as in Combination Product A. For Combination Product B, the applicant is seeking an indication of prevention and relief of bronchospasm in patients 4 years of age and older with reversible obstructive airway disease.

486

487 The same route of administration applies to Combination Products A and B. Both combination products require the same actuation force to administer an inhalation. Both combination products are intended for use in an emergency-use scenario to rapidly reverse bronchospasm.

489

491 Using the stepwise framework, the applicant's gap analysis identifies the following:

492

## ***Contains Nonbinding Recommendations***

*Draft—Not for Implementation*

493 **Step 1.** The applicant identifies the differences between Combination Products A and B. The  
494 key differences are the change in the drug constituent part and the inclusion of a pediatric age  
495 group.

496  
497 In considering the potential effect of the individual and aggregate differences in the safety  
498 and effectiveness profile of the two different drugs and the combination product as a whole,  
499 the applicant determines that full characterization of the NME will be required to establish  
500 safety and effectiveness of the drug constituent part. Additionally, the inclusion of the  
501 pediatric age group leaves an unanswered question of whether this user population can use  
502 the product safely and effectively.

503  
504 **Step 2.** The applicant has not yet developed any information for Combination Product B and,  
505 therefore, will have to either leverage existing information on the device constituent part or  
506 develop new data.

507  
508 **Step 3.** The applicant determines that the user interface is the same between Combination  
509 Products A and B, and the uses, and environments of use of the products is unchanged.<sup>19</sup> For  
510 the adult population the results of their use-related risk analysis did not identify any new or  
511 differing use-related risks between Combination Products A and B, thereby creating a bridge  
512 for adult user interface information between the products. The applicant believes, however,  
513 that an assessment will be needed in the pediatric population to assess HF since that group  
514 was not studied for Combination Product A. In addition, the applicant determines that the  
515 design control system developed for Combination Product A may be usable, subject to the  
516 assessment discussed in Step 5, for design verification and validation of Combination  
517 Product B.

518  
519 The applicant also determines that it could rely on previously conducted biocompatibility  
520 studies with Combination Product A (assessing contact of the mouth and lips with the plastic  
521 of the inhaler) because the materials remain the same.

522  
523 **Step 4.** The applicant determines that there is no other existing information that may be  
524 leveraged.

525  
526 **Step 5.** The applicant determines that it will need to conduct studies to fully characterize the  
527 NME in Combination Product B and reassess the applicability of the design inputs and  
528 outputs (design specs) for Combination Product A because of the change in drug and  
529 intended patient population. Also, because of these differences between Combination  
530 Products A and B that could affect device design and performance, the applicant determined  
531 that phase 3 clinical studies of Combination Product B, including the TBM device, are  
532 needed as well as other design verification testing for Combination Product B. In addition,  
533 the applicant intends to produce a HF validation study report for the pediatric population.

534

---

<sup>19</sup> See the guidance for industry *Safety Considerations for Product Design to Minimize Medication Errors* (April 2016) for more information.